



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/694,846	10/29/2003	Markus Hchli	8932-802-999	2490

20582 7590 04/04/2007
JONES DAY
222 East 41st Street
New York, NY 10017-6702

EXAMINER

WOODALL, NICHOLAS W

ART UNIT	PAPER NUMBER
----------	--------------

3733

SHORTENED STATUTORY PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE
3 MONTHS	04/04/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

Office Action Summary	Application No. 10/694,846	Applicant(s) HEHLI ET AL.	
	Examiner Nicholas Woodall	Art Unit 3733	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 28 December 2006.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-18 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-18 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 28 December 2006 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☒ None of:
- ☒ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

1. This action is in response to applicant's amendment received on 02/05/2007.

Drawings

2. The drawings were received on 12/28/2006. These drawings are acceptable.

Claim Rejections - 35 USC § 102

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

2. Claims 1, 2, 7, 9, 13, 14, and 17 are rejected under 35 U.S.C. 102(b) as being anticipated by Schmieding (U.S. Patent 5,626,613).

Regarding claim 1, Schmieding discloses a device having a longitudinal shape with a central axis, a first end, and a second end. The shape of the device is helical. Regarding claim 2, Schmieding discloses a device with an envelope of the helix is a cylinder having the same axis as the device. Regarding claim 7, Schmieding discloses a device wherein the cross-section orthogonal to the central axis of the helix is circular. Regarding claim 9, Schmieding discloses a device wherein the second end of the device is pointed. Regarding claim 13, Schmieding discloses a device wherein the portion nearest the first end of the device is thicker than the portion nearest the second end. Regarding claim 14, Schmieding discloses a device wherein the central axis of the helix is a straight line. Regarding claim 17, Schmieding discloses a device wherein the device is provided with through holes.

Claim Rejections - 35 USC § 103

3. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

4. Claims 8, 10, and 12 are rejected under 35 U.S.C. 103(a) as being unpatentable over Schmieding (U.S. Patent 5,626,613).

Schmieding discloses the invention as claimed except for the cross-section orthogonal to the central axis having a squared, star, rectangular, or rectangular with rounded edges shape. It would have been an obvious matter of design choice to one skilled in the art at the time the invention was made to construct the cross-section orthogonal to the central axis of the device of Schmieding having a squared, star, rectangular, or rectangular with rounded edges shape, since applicant has not disclosed that such solve any stated problem or is anything more than one of numerous shapes or configurations a person ordinary skill in the art would find obvious for the purpose of providing a forming edge in the heating portion or clamp. In re Dailey and Eilers, 149 USPQ 47 (1966).

5. Claims 3-6, 11, and 15-16 are rejected under 35 U.S.C. 103(a) as being unpatentable over Schmieding (U.S. Patent 5,626,613).

Schmieding discloses the invention as claimed except for the helix having a rotation of less than 540 degrees, the radius of the cylinder is in the range of 10 to 50 mm, etc., it would have been obvious to one having ordinary skill in the art at the time

Art Unit: 3733

the invention was made to construct the device of Schmieding with the helix having a rotation of less than 540 degrees, the radius of the cylinder being in the range of 10 to 50 mm, etc., since it has been held that where the general conditions of a claim are disclosed in the prior art, discovering the optimum or workable ranges involves only routine skill in the art. *In re Aller*, 105 USPQ 233.

6. Claim 18 is rejected under 35 U.S.C. 103(a) as being unpatentable over Schmieding (U.S. Patent 5,626,613).

Schmieding discloses the invention as claimed except for the device having at least two through holes. It would have been obvious to one having ordinary skill in the art at the time the invention was made to construct the device of Schmieding having at least two through holes, since it has been held that mere duplication of the essential working parts of a device involves only routine skill in the art. *St. Regis Paper Co. v. Bemis Co.*, 193 USPQ 8.

Response to Arguments

Applicant's arguments filed 12/28/2006 have been fully considered but they are not persuasive. The applicant's argument that Schmieding does not disclose an intramedullary nail that is inserted into the intramedullary canal of a bone is not persuasive. The examiner would like to point out that the statements of intended use and other functional statements do not impose any structural limitations on the claims distinguishable over Schmieding, which is capable of being used as claimed if one so desires to do so. *In re Casey*, 152 USPQ 235 (CCPA 1967) and *In re Otto*, 136 USPQ 458, 459 (CCPA 1963). Furthermore, the law of anticipation does not require that the

reference "teach" what the subject patent teaches, but rather it is only necessary that the claims under attack "read on" something in the reference. *Kalman v. Kimberly Clark Corp.*, 218 USPQ 781 (CCPA 1983). Furthermore, the manner in which a device is intended to be employed does not differentiate the claimed apparatus from prior art apparatus satisfying the claimed structural limitations. *Ex parte Masham*, 2 USPQ2d 1647 (1987). The examiner believes the device of Schmieding is capable of being used as an intramedullary nail in the bone of an organism smaller than a human.

Conclusion

7. **THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Nicholas Woodall whose telephone number is 571-272-5204. The examiner can normally be reached on Monday to Friday 8:00 to 5:30 EST..

Art Unit: 3733

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on 571-272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

NWW


EDUARDO C. ROBERT
SUPERVISORY PATENT EXAMINER